

## The Annual QC Result of FUJI DRI-CHEM 3500 (2012)

ITEM	Unit	Mean	QP-L		Mean	QP-H	
			Expected Range	Detected Range		Expected Range	Detected Range
GLU	mg/dl	106	95-116	96	306	277-335	276
BUN	mg/dl	13.8	11.9-15.6	13.6	49.9	45.0-54.8	48.6
UA	mg/dl	3.8	3.2-4.3	3.9	11.0	10.1-11.9	10.9
TCHO	mg/dl	109	96-122	113	145	129-161	149
TG	mg/dl	122	103-141	130	217	193-242	223
CRE	mg/dl	1.0	0.5-1.4	0.9	5.1	4.5-5.8	5.0
TP	g/dl	5.8	5.2-6.4	5.6	7.2	6.6-7.8	7.1
ALB	g/dl	3.5	2.9-4.1	3.9	4.2	3.6-4.8	4.5
TBIL	mg/dl	0.9	0.6-1.2	1.1	4.6	3.8-5.5	5.1
HDL	mg/dl	33	23-42	33	***	***	***
Ca	mg/dl	9.5	8.2-10.8	9.4	13.4	11.9-15.0	12.9
IP	mg/dl	3.5	3.0-3.9	3.6	6.1	5.4-6.9	6.3
Mg	mg/dl	0.4	0.1-0.7	0.4	***	***	***
GOT	U/l	31	25-37	29	99	87-110	91
GPT	U/l	38	32-44	37	98	87-110	104
CPK	U/l	77	60-93	78	302	262-343	309
LDH	U/l	159	138-180	152	419	362-476	418
ALP	U/l	150	118-183	119	456	386-527	372
AMYL	U/l	49	37-61	48	173	148-198	193

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator: 黃雅倫

Date: 3/19/2012

Pathology  
Core Director: 陳國強

Date: 3/20/2012

Executive  
Assistant: Jo-Ping Sun

Date: 3/21/2012