

The Annual QC Result of FUJI DRI-CHEM 4000 (2015)

		QP-L			QP-H		
ITEM	Unit	Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	107	96-118	112.67	297	269-325	313.33
BUN	mg/dl	13.8	12.0-15.7	14.23	49.8	45.1-54.5	51.10
UA	mg/dl	3.8	3.2-4.3	3.97	11.0	10.1-11.8	11.60
TCHO	mg/dl	111	98-124	120.00	144	128-160	154.67
TG	mg/dl	132	113-152	139.67	228	203-253	246.00
CRE	mg/dl	1.0	0.6-1.4	1.03	5.2	4.6-5.7	5.47
TP	g/dl	5.7	5.1-6.3	5.87	7.1	6.5-7.7	7.27
ALB	g/dl	3.4	2.8-4.0	3.93	4.2	3.6-4.8	4.60
TBIL	mg/dl	0.9	0.6-1.2	1.10	4.9	4.1-5.7	5.40
HDL	mg/dl	30	21-39	30.00	**	***	**
Ca	mg/dl	9.9	8.5-11.2	10.57	14.0	12.5-15.6	15.07
IP	mg/dl	3.5	3.0-3.9	3.83	6.3	5.6-6.9	6.70
Mg	mg/dl	1.4	1.1-1.7	1.50	**	***	**
GOT	U/l	33	27-39	31.00	97	87-108	95.00
GPT	U/l	39	33-45	38.00	93	82-104	94.33
CPK	U/l	88	72-104	91.67	313	271-355	331.67
LDH	U/l	170	147-192	178.33	440	382-497	468.33
ALP	U/l	164	132-197	169.33	477	413-541	455.67
AMYL	U/l	51	39-63	50.66	170	150-190	167.67

The data result of each item was repeated three times.

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator: 林郁芳

Date: 2015年 4月 24日

Pathology
Core Director: _____



Date: _____

TMC Director: 陳志強

Date: 2015, 4. 27