

The Annual QC Result of FUJI DRI-CHEM 4000 (2020)

		QP-L			QP-H		
ITEM	Unit	Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	105	93-117	106.67	313	283-343	321.33
BUN	mg/dl	14.0	12.1-15.8	13.97	50.1	45.2-55.0	50.77
UA	mg/dl	3.6	3.0-4.1	3.70	10.7	9.8-11.6	11.13
TCHO	mg/dl	120	107-134	126.67	139	124-154	147.00
TG	mg/dl	132	112-151	133.33	232	207-257	240.67
CRE	mg/dl	0.99	0.60-1.38	1.01	5.22	4.62-5.82	5.54
TP	g/dl	5.7	5.1-6.3	5.70	7.0	6.4-7.6	6.97
ALB	g/dl	3.6	3.0-4.2	3.53	4.6	4.0-5.2	4.50
TBIL	mg/dl	0.9	0.6-1.1	1.00	4.4	3.6-5.2	4.50
HDL	mg/dl	36	26-45	36.00	**	***	**
Ca	mg/dl	9.1	7.8-10.4	8.00	12.7	11.1-14.2	11.43
IP	mg/dl	3.6	3.2-4.1	3.83	6.4	5.7-7.0	6.77
Mg	mg/dl	1.4	1.1-1.7	1.30	**	***	**
GOT	U/l	32	25-38	32.00	90	80-101	101.00
GPT	U/l	36	30-43	36.33	92	81-103	93.00
CPK	U/l	68	52-84	69.67	246	213-280	264.33
LDH	U/l	160	133-186	155.67	397	343-452	400.00
ALP	U/l	180	147-212	178.33	553	477-629	542.33
AMYL	U/l	56	44-69	53.33	184	158-210	170.67

The data result of each item was repeated three times.

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator:

紀宣仔

Date:

2020.7.13

Pathology

Core Director:

張超坤

Date:

20.20.7.13

TMC Director:

陳志文

Date:

2020.7.13