


The Annual QC Result of Fuji DRI-CHEM

ITEM	UNIT	QP-L		QP-H	
		Expected Range	Detected Value	Expected Range	Detected Value
GLU	mg/dl	96-118	100	277-335	292
BUN	mg/dl	11.6-15.2	14.4	45.0-54.8	49.5
UA	mg/dl	3.3-4.3	3.8	10.1-11.9	11.4
TCHO	mg/dl	91-115	94	129-161	153
TG	mg/dl	126-159	150	208-258	252
CRE	mg/dl	0.5-1.4	0.9	4.5-5.8	5.1
TP	g/dl	5.2-6.4	5.7	6.6-7.8	7.1
ALB	g/dl	2.7-3.9	3.5	3.6-4.8	4.3
TBIL	mg/dl	0.8-1.4	1.3	3.8-5.5	4.9
HDL	mg/dl	17-29	25	--	--
Ca	mg/dl	6.3-9.0	7.6	11.9-15.0	13.2
IP	mg/dl	3.0-3.9	3.5	5.4-6.9	5.9
Mg	mg/dl	0.2-0.8	0.4	--	--
GOT	U/l	28-40	29	87-110	90
GPT	U/l	36-46	34	87-110	93
CPK	U/l	48-80	60	262-343	333
LDH	U/l	131-172	143	362-476	392
ALP	U/l	115-179	124	386-527	403
AMYL	U/l	37-61	47	148-198	168

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator: 黃雅倫

Date: 2011/03/04

Pathology
Core Director: 

Date: _____

TMC Director: 

Date: 2011/3/4