

## The Annual QC Result of FUJI DRI-CHEM 4000 (2022)

		QP-L			QP-H		
ITEM	Unit	Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	108	95-120	113.67	321	290-352	341.33
BUN	mg/dl	14.0	12.1-15.8	14.57	50.6	45.7-55.6	55.07
UA	mg/dl	3.7	3.1-4.2	4.13	11.0	10.1-11.9	11.80
TCHO	mg/dl	123	109-137	130.33	154	137-171	168.67
TG	mg/dl	134	114-154	143.33	238	213-263	256.00
CRE	mg/dl	0.98	0.59-1.37	1.03	5.16	4.56-5.76	5.74
TP	g/dl	5.7	5.1-6.3	5.90	7.2	6.5-7.8	7.50
ALB	g/dl	3.5	2.9-4.1	3.57	4.5	3.9-5.1	4.40
TBIL	mg/dl	0.6	0.4-0.9	0.87	3.9	3.1-4.7	4.60
HDL	mg/dl	37	27-46	40.67	**	***	**
Ca	mg/dl	9.5	8.2-10.8	8.40	13.6	12.0-15.2	13.27
IP	mg/dl	3.7	3.2-4.1	4.03	6.5	5.8-7.2	7.17
Mg	mg/dl	1.4	1.1-1.7	1.33	**	***	**
GOT	U/l	33	27-39	34.67	92	81-102	101.00
GPT	U/l	46	40-53	49.00	109	96-121	115.33
CPK	U/l	76	60-92	81.33	268	231-304	292.67
LDH	U/l	157	130-183	172.33	444	383-504	469.00
ALP	U/l	188	156-221	162.67	560	483-637	515.33
AMYL	U/l	53	41-66	52.00	175	150-201	170.67

The data result of each item was repeated three times.

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator:

蔡宜芬

Date:

2022.03.31

Pathology

陈昱彰

Date:

2022.04.06

TMC Director:

赖超坤

Date:

2022.04.08