

The Annual QC Result of FUJI DRI-CHEM 4000 (2024)

ITEM	Unit	QP-L			QP-H		
		Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	108	95-120	108.0	315	285-345	320.3
BUN	mg/dl	14.5	12.6-16.3	14.5	51.4	46.4-56.4	52.8
UA	mg/dl	3.5	2.9-4.0	3.4	10.5	9.6-11.3	10.6
TCHO	mg/dl	135	120-150	134.7	167	149-185	175.3
TG	mg/dl	129	109-148	121.7	223	199-247	218.3
CRE	mg/dl	0.99	0.60-1.38	0.96	5.42	4.80-6.04	5.7
TP	g/dl	5.9	5.3-6.5	6.0	7.2	6.6-7.9	7.3
ALB	g/dl	3.6	2.9-4.2	3.9	4.7	4.1-5.4	4.9
TBIL	mg/dl	0.8	0.5-1.1	0.7	4.0	3.2-4.8	3.9
HDL	mg/dl	42	34-50	44.0	**	***	**
Ca	mg/dl	8.6	7.3-9.9	8.4	12.4	10.8-13.9	12.1
IP	mg/dl	3.6	3.1-4.0	3.8	6.2	5.6-6.9	6.2
Mg	mg/dl	1.6	1.3-1.9	1.5	**	***	**
GOT	U/l	32	26-38	32.3	95	85-106	92.0
GPT	U/l	38	32-44	39.3	106	94-119	105.0
CPK	U/l	68	52-84	71.0	283	244-321	282.3
LDH	U/l	171	143-199	169.7	476	412-540	518.0
ALP	U/l	187	154-219	177.0	615	531-699	594.0
AMYL	U/l	61	49-74	59.3	205	177-232	203.7

The data result of each item was repeated three times.

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator: 蔡宜芬 Date: 2024/04/02

TMC Pathology
Operating Officer: 陳昱廷 Date: 2024/04/02

TMC Chief Operating
Officer: 賴超坤 Date: 2024/04/02

TMC
Program Director: 陳志威 Date: 2024/04/03