


# The Annual QC Result of FUJI DRI-CHEM 4000 (2025)

		QP-L			QP-H		
ITEM	Unit	Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	108	95~120	113.3	319	288~349	330.0
BUN	mg/dl	14.0	12.1~15.9	14.2	51.3	46.3~56.3	53.6
CRE	mg/dl	0.97	0.58~1.36	1.0	5.43	4.81~6.05	5.6
ALB	g/dl	3.6	2.9~4.2	3.8	4.6	4.0~5.3	5.0
TBIL	mg/dl	0.7	0.4~1.0	0.8	4.1	3.3~5.0	4.8
GOT	U/l	37	30~43	38.0	90	80~100	92.8
GPT	U/l	41	35~47	43.3	101	89~112	101.8
ALP	U/l	207	172~242	185.5	626	541~712	640.3


The data result of each item was repeated three times.  
This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator:




Date: 2025/02/19

TMC Pathology  
Operating Officer:




Date: 2025/02/19

TMC Chief Operating  
Officer:



Date: 2025/02/19

TMC  
Program Director:



Date: 2025/02/19