

The Annual QC Result of FUJI DRI-CHEM 4000 (2026)

ITEM	Unit	QP-L			QP-H		
		Mean	Expected Range	Detected Value	Mean	Expected Range	Detected Value
GLU	mg/dl	108	95~120	105	319	288~349	316
CRE	mg/dl	0.97	0.58~1.36	0.94	5.43	4.81~6.05	5.45
TBIL	mg/dl	0.7	0.4~1.0	0.7	4.1	3.3~5.0	4.2
GOT	U/l	37	30~43	34	90	80~100	86
GPT	U/l	41	35~47	40	101	89~112	90

The data result of each item was repeated three times.

This table showed that all the detected values of different items are fall within the expected range provided by FUJI.

Operator:

蔡宜芬

Date:

2026/03/20

TMC Pathology

Operating Officer:

陳昱豪

Date:

2026/03/23

TMC Chief Operating
Officer:

賴超坤

Date:

2026/03/23

TMC

Program Director:

陳志和

Date:

2026/03/23